

SERVOMEX 2200 SERIES

PROCESS OXYGEN ANALYSERS

for use on demanding process control applications, including safety critical, purity, feedstock clean-up measurements where high performance is critical

- Unrivalled accuracy, reliability and repeatability
- Low cost of long term ownership
- Up to six transmitters can be connected to a single control unit
- Hazardous area ATEX Zone 1, FM, CSA, TIIS, IECEx
- Transmitter interface software available




SPECIFICATION

	TRANSMITTER	TRANSMITTER	TRANSMITTER	CONTROL UNIT	CONTROL UNIT
TECHNOLOGY	Paramagnetic	Paramagnetic	Paramagnetic	-	-
Sample Gas Temperature :	60°C	<110°C *	<135°C *		
Hazardous area:	ATEX Cat. 2 IECEX FM Class I Div. 1 CSA TIIS	ATEX Cat.2 IECEX FM Class I Div. 1 CSA	ATEX Cat.2 IECEX FM Class I Div. 1 CSA	ATEX Cat.2 IECEX FM Class I Div.1 CSA TIIS	ATEX Cat.3 IECEX FM Class I Div.2 CSA
PERFORMANCE					
Maximum range:	0-100%	0-100%	0-100%		
Minimum range:	0-0.5%	0-1%	0-1%		
Intrinsic Error (accuracy):	<0.02% O ₂	<0.04% O ₂	<0.08% O ₂		
Linearity:	<0.01% O ₂	<0.02% O ₂	<0.04% O ₂		
Repeatability:	0.02%	0.03%	0.04%		
Response time (T ₉₀):	<4 seconds	<4 seconds	<4 seconds		
Zero drift per week:	<0.02% O ₂ /week	<0.04% O ₂ /week	<0.08% O ₂ /week		
Span drift per week:	<0.05% O ₂ /week	<0.10% O ₂ /week	<0.20% O ₂ /week		


* Temperature of operation: Consult Servomex for other temperatures

Hazardous Area Approvals


2210 Control Unit

ATEX  II 3(I)GD EEx nCL ia IIC T4
FM Class I, Division 2
CSA Class I, Division 2
IEC Ex


2213 Control Unit

ATEX  II 2(I)GD EEx d ia IIC T4
FM Class I, Division 1
CSA Class I
TIIS
IEC Ex

2223 Transmitter

ATEX  II 2(I)GD EEx iad IIC T4
FM Class I, Division 1
CSA Class I
TIIS
IEC x

2222 Transmitter

ATEX  II 2(I)GD EEx iad IIC T3
FM Class I, Division 1
CSA Class I
IEC Ex

EC Directive Compliance

2210/2213 Control Units & 2223 Transmitters

comply with the Low Voltage Directive 73/23/EEC and the EMC Directive 89/336/EEC (as amended by Directive 92/31/EEC), both as amended by Directive 93/68/EEC.

They conform to the following harmonised European standards for product safety and electromagnetic compatibility:

EN 61010-1: Safety requirements for electrical equipment for measurement, control and laboratory use.

EN 61326-1: Electrical equipment for measurement, control and laboratory use - EMC requirements (All induced errors are less than the intrinsic error).

These products are rated for "Installation Category II" in accordance with IEC 664.

These products are rated for "Pollution Degree 2" in accordance with IEC 664.

EC Directive Compliance

2200H 110°C & 135°C (2222) Transmitters

comply with the EMC Directive 89/336/EEC (as amended by Directive 92/31/EEC), as amended by Directive 93/68/EEC.

They conform to the following harmonised European standards for product safety and electromagnetic compatibility:

EN 61010-1: Safety requirements for electrical equipment for measurement, control and laboratory use.

EN 61326 1997 + A1: 1998 + A2: 2001: Electrical equipment for measurement, control and laboratory use - EMC requirements See manual for further information.

These products are rated for "Installation Category II" in accordance with IEC 664.

These products are rated for "Pollution Degree 2" in accordance with IEC 664.

SERVOMEX COMPANIES, AGENTS AND REPRESENTATIVES ARE LOCATED THROUGHOUT THE WORLD. FOR YOUR NEAREST CONTACT VISIT WWW.SERVOMEX.COM

Servomex has a policy of constant product improvement and therefore reserves the right to change specifications without notice.



Certificate No. Q05166
BS EN ISO 9001



Certificate No. EMS82701
BS EN ISO 14001



EAME (Europe, Africa and Middle East):

Servomex Group Limited, Jarvis Brook, Crowborough, East Sussex, TN6 3DU, England

Servomex GmbH, Münsterstraße 5, 59065 Hamm, Germany

Servomex S.A., 8 Rue Proudhon, B.P. 50, 93212 St Denis La Plaine Cedex, France

Servomex B.V., Stephensonstraat 20, 2723 RN Zoetermeer, Netherlands

Servomex Group Limited, SAIF Zone, Executive Suites Z1-56, PO Box 8962, Sharjah, UAE.

AMERICAS:

Servomex Company Inc., 525 Julie Rivers, Suite 185, Sugar Land, Texas 77478-2847, USA.

ASIA PACIFIC:

Servomex Asia Pacific Ltd, (Spectris Instrumentation and Systems (Shanghai) Ltd)

Building 9, Shanghai Euromate Industrial Park, No.88 Lane 2888 Hua Ning Road, Xinzhuang Industrial Park, Shanghai 201108 PR China

Servomex Asia Pacific Ltd, 7F, No.88, Sec. 6, Chung Shan N. Rd., Taipei, Taiwan

Servomex India, (Spectris Technologies Pvt.Ltd) 7th Floor, Sunil Enclave, CTS No.307, Village Gundavli, Shivaji & Perriera Hill Road Junction,

Opp. Landmark Building, Off Andheri Kurla Road, Andheri (East), Mumbai 400 099, India.

Servomex Japan, (Spectris Co.Ltd) 1F JP Kaji-cho Bldg., 3-5-2 Kandakaji-cho, Chiyoda-ku, Tokyo 101-0045, Japan

Global email: info@servomex.com

☎ (44) 1892 652181 Fax (44) 1892 662253

☎ (49) 23 81 68 82 13 Fax (49) 23 81 68 81 75

☎ (33) 1 49 46 22 50 Fax (33) 1 49 46 22 51

☎ (31) 79-346 42 42 Fax (31) 79-342 08 19

☎ (971) 6 557 0303 Fax (971) 6 557 1242

☎ (1) 281-295-5800 Fax (1) 281-295-5899

☎ (86) 21 6442 6488 Fax (86) 21 6442 6498

☎ (886) 2-2833 8848 Fax (886) 2-2833 8844

☎ (91) 22 67100359/60 Fax (91) 22 56778896

☎ (81) 3 3255 8162 Fax (81) 3 3255 8154

Website: <http://www.servomex.com>